

The Philips logo, consisting of the word "PHILIPS" in a bold, blue, sans-serif font.The Capsule logo, featuring the word "capsule" in a lowercase, blue, sans-serif font, followed by a red circular icon containing a white stylized 'e'.

# Capsule Vitals Plus Regulatory Notice on Neuron 3

## Introduction

---

Capsule Vitals Plus is a Patient Monitoring system which includes hardware and software, as well as a centralized configuration, for the collection, management, and transmission of clinical information, in both critical and non-critical care environments.

### ***This Document***

This Notice contains regulatory information for the system and is therefore an extension of the *Capsule Vitals Plus Installation and Maintenance Guide*. It details the compliance statements that the product requires for its certification and approval. Philips is committed to delivering products compliant with standards, laws and regulations.

## Certification Marks



Recognized component



Separate collection for electrical and electronic waste (WEEE directive and Battery directive)



NRTL mark for USA and Canada



Global Trade Item Number

**Rx Only**

Federal law restricts this device to sale by or on the order of a physician (USA audiences only)



Unique Device Identifier

## Regulatory compliance and approvals

### USA

#### Federal Communications Commission

Capsule Vitals Plus complies with Part 15 of the FCC Rules. Operation is subject to two conditions:

- This device may not cause harmful interference.
- This device must accept any interference received, including interference that may cause undesired operation.

**Note:** This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help

**Note:** Vitals Plus systems that integrate any revision of the Capsule Neuron 3 are Class-B compliant.

**Note:** For operation within 5.15 ~ 5.25GHz / 5.47 ~5.725 GHz frequency range, Capsule Vitals Plus is restricted to indoor environment. This device meets the requirements specified in Part 15E, Section 15.407 of the FCC Rules.

**Caution:** Philips is not responsible for any radio or television interference caused by unauthorized changes or modifications to this equipment. Unauthorized changes or modifications could void the user's authority to operate the equipment.



**Caution:** This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter. Unauthorized changes or modifications could void the user's authority to operate the equipment.

#### Radiation Exposure Statement

This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance 20 cm between the radiator and your body.

**Note:** The country code selection is for non-US model only and is not available on all US models. Per FCC regulation, all Wi-Fi products marketed in the USA must be fixed to US-operation channels only.

## Disposal

---

As you use the Vitals Plus system, you will accumulate solid wastes that require proper disposal or recycling. These include system components, batteries, patient applied parts, and packaging materials.

### ***Recycling and the Environment***

Improper disposal of IT and medical equipment can have a negative impact on health and the environment. We recommend that you dispose of Philips products, including all electronic devices, cables, batteries, and so on, at an appropriate facility to enable recovery and recycling. You can also recycle packaging and guides according to your local recycling regulations.

### ***Vitals Plus components***

At the end of its service life, the products described in this manual, as well as their accessories, must be disposed of in compliance with the guidelines regulating the disposal of such products. If you have questions concerning disposal of the product, contact your Philips representative.

### ***Patient Applied Parts***

The following patient applied parts are intended for single use, and should be disposed of properly as medical waste in accordance with regional body-controlled guidelines and hospital protocol:

- Single-use blood pressure cuffs
- Single-use SpO<sub>2</sub> sensors
- Temperature probes
- Scanner sheaths
- Reusable caps

Other patient applied parts are intended for single-patient use, and should similarly be disposed of properly as medical waste in accordance with regional body-controlled guidelines and hospital protocol:

- Sampling lines

Other patient applied parts should be cleaned according to manufacturer instructions. Inspect reusable applied parts for wear, replace as necessary, and dispose of used product as medical waste in accordance with regional body-controlled guidelines and hospital protocol:

- Multi-patient reusable cuffs
- Reusable SpO<sub>2</sub> sensors

## **Batteries**

The sealed, rechargeable batteries contain lead and can be recycled. Discharge (deplete) batteries prior to disposal. Place the battery in packaging which electrically isolates its contents. Do not puncture or place the battery in a trash compactor. Do not incinerate the battery or expose it to fire or high temperatures. Dispose in accordance with regional body-controlled guidelines and hospital protocol.

**Warning:** Improper disposal of batteries may create an explosion or contamination hazard. Always recycle batteries according to local regulations. Never dispose of batteries in refuse containers. Do not wrap them in metal or aluminum foil. Wrap them in newspaper before disposing of them. Do not burn them. Battery may explode if overheated.

## **Packaging Material**

Retain original packaging materials for future use in shipping the system and its accessories. The recommendation includes corrugated shippers and inserts. Whenever possible, recycle the packaging of accessories and patient applied parts.

## Standards and Regulations

Field	Standard or regulation
Medical device safety	IEC 60601-1 IEC 60601-1-2 IEC 80601-2-30 (NIBP) ISO 81060-2 (NIBP) ISO 80601-2-56 (Temperature) ISO 80601-2-61 (Pulse oximeter) IEC 80601-2-49 (Patient monitoring) IEC 60601-1-8 (Alarms) IEC 60601-2-55 (Capnography)
Medical device usability	IEC 60601-1-6 IEC 62366-1 IEC 62366
Medical device Software – Software Lifecycle Processes	IEC 62304
EMC/EMI	FCC 47 CFR Part 15 sub-part B CAN ICES-003 B / NMB-003 B IEC 60601-1-2
Radio	SmartLinx Neuron 3 complies with <ul style="list-style-type: none"> <li>• FCC 47 CFR Part 15 subparts C and E</li> <li>• RSS-210</li> <li>• RSS-247</li> <li>• ANSI C63.27</li> </ul>
RF exposure	ANSI/IEEE C95.1 RSS-102

**Legal Notice**

© Koninklijke Philips N.V. and/or its subsidiaries 2015. All rights reserved.

This product and related documentation are protected by copyright and distributed under licensing restricting their use, copying, distribution, and decompilation. No part of this product or related documentation may be reproduced in any form or by any means without prior written authorization of Koninklijke Philips N.V. and/or its subsidiaries.

This technical data may be subject to U.S. and international export, re-export, or transfer ("export") laws. Diversion contrary to U.S. and international law is strictly prohibited.

The Capsule logo is a registered trademark of Koninklijke Philips N.V. and/or its subsidiaries in the US, EU and other countries. All other trademarks, service marks, registered trademarks and registered service marks are the property of their respective owners.

The hardware and software of Koninklijke Philips N.V. and/or its subsidiaries is provided subject to all third-party licenses and limitations in the "third party notices" electronic file included as part of the software or available upon request from Koninklijke Philips N.V. and/or its subsidiaries.

DHF21705

DCN 2023-054